



Attestation of Conformity

ICR Polska



Product name: Disposable medical face mask
Product type/model: A1, A2, A3
Trade mark: n/a

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

Relevant EC Directive: Medical Device Directive 93/42/EEC
Conformity assessment procedure: EC Declaration of Conformity (Annex VII of Directive 93/42/EEC)
Classification: Class I according Rule 1 of Annex IX of Directive 93/42/EEC
Applied normative documents: EN 14683+2005
Applied Quality Management System: n/a

This AoC will remain valid only if Quality Management System Certificate remains valid. The assessment process has been carried out in accordance with the program PC-P-07-07. Evaluation has been carried out in accordance with test report made by:

- Shenzhen MONLKA Technology Co., Ltd

No. of test report: MNK20200311043R
Issue date: 16.03.2020
Expiration date: 15.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR Polska/2020-2040.

This Attestation applies to products having the same attributes (parameters), intended use, that have been evaluated and meet the requirements of the aforementioned standard.



Director: Rafał Kalinowski

Warsaw, 16. 03. 2020.

ICR Polska Co. Ltd.
 ul. Plac Przymierza 6, 03-944 Warszawa
 www.icrpolska.com, e-mail: icrpolska@icrqa.com